

GOVERNMENT OF WEST BENGAL
OFFICE OF THE DIRECTOR
Institute OF Post Graduate Medical Education & Research
244 , A.J.C. Bose Road , Kolkata - 700 020

Memo No. Inst/ IPGMER/ 10392

Dated : 04-10-18

NOTICE

Sealed quotations are invited from bonafide/interested firms with superscription " Quotation for Embalming Machine (Cadaver Injector) vide memo no. Dated..... ". The quotation should reach the office of the undersigned on or before **11/10/2018 by 3 pm** . The quotation should accompany current valid trade Licence, PAN No., GST Registration Proof & **Product / Equipments catalogue for specification proof(if any) .**

Quotation will Valid upto 1 Yrs. from the date of approval .

The basic rate should be furnished inclusive Transportation Cost, Insurance, Testing & Incidental Charges including others (if any) of the instrument etc. but excluding of GST, which noted separate column as mentioned below. **The warranty of the instrument 1yrs minimum after instalation.**

Acceptance of the lowest rate is not obligatory for the undersigned who reserves the right to accept or reject any offer without assigning any reason whatsoever.

The Quotation may be opened on 12/10/2018 onward, Bonafide/interested firm's representative may be present in time of opening at the office of undersigned .

Sl.no.	Name of the Articles	Rate/Unit	GST	Total	Mfg & Model no
1	Embalming Machine (Cadaver Injector)				

Product Specification :

- i) For injecting formaldehyde solution in Cadaverous at much higher speed than normal gravity process .
- ii) Unit is fully covered & mounted on a portable trolley having four castor wheels for easy movement .
- iii) Unit consists of one air compressor fitted with ½ hp to 1 hp motor which is connected with a stainless steel tank of 10 litres capacity meant for storing and injecting the solution.
- iv) Tank is fitted with a safety valve, pressure gauge and rubber tubing having provision for injection .
- v) Supplied complete with electric cord, plug and others suitable to work on 220 volts, 1 ph 50 Hz , AC supply .


4/10/18
Director