

GOVERNMENT OF WEST BENGAL

OFFICE OF THE DIRECTOR

Date 26-8-18

MEMONO - 9400


IPGME & R, KOLKATA  
NOTICE

Sealed tenders are invited from bona fide/interested firms with superscription "Tender for WBDST funded project entitled "Studies on Genetic Polymorphism of RET Proto oncogene in Hirschsprung Disease in Children" (MEMO No.:154(Sanc.)/ST/P/S&T/9G-15/2016) dt. 08/02/2018 in Dept. of Pathology, IPGME & R". The tenders should reach the office of the undersigned on or before 05/09/18. The tenders should accompany with current valid trade license, PAN No. and GST registration certificate proof. Acceptance of the lowest rate is not obligatory for the undersigned who receives the right to accept or reject any offer without assigning any reason whatsoever.

Rate must be submitted with Company Name and Cat. No. after calculating GST or other taxes, if any, according to the following format. The tender may be opened on 06/09/18 onward; bona fide/interested firms' representatives may contact.

NAME OF THE ITEM	SPECIFICATION	UNIT
Fluorimeter	1) Linear Dynamic Range - Five (5) orders of magnitude, assay-dependent 2) Detectors present - Solid state silicon sensor 3) Excitation Peak - Red: 627nm Blue: 470nm 4) Excitation Filters - Red: 640nm shortpass (wavelengths up to 640nm) Blue: 495nm shortpass (wavelengths up to 495nm) 5) Emission Filters - Red: 660-720nm Blue: 510-580nm 6) Sample Tube - 0.5ml PCR tube 7) Type of Calibration - Single point 8) Power Requirements for the instrument - 5V, 0.2A Maximum 9) Automatic Sleep Mode to be present - After 1 hour of inactivity 10) Operating System Requirements (for optional Quantus™ Software) - Windows® 8, Windows® 7 (32 or 64 bit) .NET Framework 4.0 or later. At least one USB port available	One

MEMONO 9400/1(A) dt. 26.8.18

  
 21/8/18  
 Director  
 I.P.G.M.E.&R  
 Kolkata-700010  
 DIRECTOR, IPGME & R.