

**GOVERNMENT OF WEST BENGAL**  
**OFFICE OF THE DIRECTOR**  
**Institute of Post Graduate Medical Education and Research**  
**244, AJC Bose Road, Kolkata-700020**

Memo No.- 8531

Date: 29/05/2018

Sealed quotations are invited from bona fide/interested farms with superscription "Quotation for WBDST Funded project in Dept. Of Microbiology, IPGME&R". The quotations should reach the office of the undersigned on or before 09/6/18. The quotations should accompany with current valid trade license, PAN No. and GST registration certificate proof. Acceptance of the lowest rate is not obligatory for the undersigned who receives the right to accept or reject any offer without assigning any reason whatsoever.

Rate must be submitted with company name and cat. no. after calculating GST or other taxes, if any, according to the following format.

The quotation may be opened on 11/6/18 onward; bona fide/interested farm's representative may be contact.

Name of the item	Specification	Unit
Adjustable Volume Single Channel Pipette(100-1000 µl)	<ul style="list-style-type: none"> <li>• ISO 8655 certified, fully Autoclavable, single channel pipette of variable volume, compatible with universal tips</li> <li>• Variable Volume Pipette- Volume: 100-1000 µl</li> <li>• Increment : 5µL</li> <li>• Accuracy : ±1.0 to 0.6%</li> <li>• Precisions: 0.6 to 0.2%</li> <li>• From reputed manufacturers.</li> </ul>	ONE
Adjustable Volume Single Channel Pipette(2-20 µl)	<ul style="list-style-type: none"> <li>•ISO 8655 certified, fully Autoclavable, single channel pipette of variable volume, compatible with universal tips.</li> <li>•Variable Volume Pipette- Volume: 2-20 µl</li> <li>•Increment : 0.1µL</li> <li>•Accuracy : ±3.0 to 1.0%</li> <li>•Precisions: 2.5 to 0.4%</li> <li>•From reputed manufacturers.</li> </ul>	ONE
Adjustable Volume Single Channel Pipette(20-200 µl)	<ul style="list-style-type: none"> <li>•ISO 8655 certified, fully Autoclavable, single channel pipette of variable volume, compatible with universal tips.</li> <li>•Variable Volume Pipette- Volume: 20-200 µl</li> <li>•Increment : 1µL</li> <li>•Accuracy : ±1.8 to 0.6%</li> <li>•Precisions: 0.7 to 0.2%</li> <li>•From reputed manufacturers.</li> </ul>	ONE

Prasanta Kumar Maity  
25/5/18

**Prof. P. K. MAITI**  
**H.O.D. Microbiology**  
**I.P.G.M.E.&R, Kol-20**

*(Signature)*  
28/5/18

Director  
IPGME&R, Kolkata

*(Signature)*  
**Director**  
**I.P.G.M.E.&R**  
**Kolkata-700020**