

GOVERNMENT OF WEST BENGAL
OFFICE OF THE DIRECTOR
Institute OF Post Graduate Medical Education & Research
244 , A.J.C. Bose Road , Kolkata - 700 020

Memo No. Inst/Store/Q/ 6339

Dated: 13.10.17

NOTICE

Sealed quotations are invited from bonafide manufacturer/authorized distributor with superscription "Quotation for Purifier for the Dept. of Microbiology". The quotation should reach the office of the undersigned on or before **24/10/2017 by 2 pm**. The quotation should accompany current valid trade Licence, PAN No., GST Registration Proof, & in case of distributor authorization letter . The authorized representative of the firm may contact the Pharmacist of the IPGME&R Store for details/specification of the items.

Acceptance of the lowest rate is not obligatory for the undersigned who reserves the right to accept or reject any offer without assigning any reason whatsoever. Rate must be submitted after calculating GST or other Taxes, if any according to the following format. The Quotation may be opened on 25/10/2017 onward, Bonafide/interested firm's representative may be contact with store.

Sl.no.	Name of the Articles	Rate/Unit including GST
1	Water Purifiers (as per specification)	per unit


13/10/17
Director

IPGME&R, Kolkata-20

Dated: 13.10.17

Technical specification:

Water Purifiers

- DIMENSIONS MM (WxDxH) 330 x 440 x 490
- Installation Type Table Top
- Storage Capacity 3.0 Litres (Cold tank - 1.5 litres & Ambient tank - 1.5 litres)
- Recommended for Water with high TDS, having a salty/brackish taste
- Applicable TDS Range 1 - 2000 mg / litre
- Water Flow Rate 15 litre/hour
- Input Water Temperature 10 - 40°C
- Membrane Type Thin Film Composite RO Membrane with UV Lamp
- Operating Input Voltage 230V / 50Hz compressor
- Water Recovery 25% approx in RO Mode

Pr
21/8/15
Prof. P. K. MAITI
H.O.D. Microbiology
I.P.G.M.E.&R, Kōl-20