



Government of West Bengal
Office of the Medical Superintendent-cum-Vice Principal
IPGME&R SSKM Hospital, 244 AJC Bose Road, Kolkata-700020
EPABX: 22041100; E-Mail: msvpsskm@gmail.com; Office: 22231589; FAX: 22234658

Memo. No. IPGME&R SSKMH/MSVP/1535/2023

Dated: 1/3/23

QUOTATION NOTICE

Sealed quotations are invited from the reputed Companies/ Direct Importers/ Authorized Distributors/ Traders for supply of the following item as mentioned below for use at the Eye Bank of IPGME&R SSKM Hospital.

The last date of submission of quotation is on 09/03/23 and the same will be opened on the same date at 2.30 P.M. in presence of interested bidders, if any. The bidders will submit the quotation in their own official pad along with the self attested copies of the following essential documents.

1. Valid Trade License/Enlistment
2. PAN Card of the Bidder/ Bidder Company
3. GST Registration Certificate

The quotation is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from from 10.00 AM to 2.00 PM.

The undersigned reserves the right to accept or cancel the whole procedure at any time without showing any reason. Rate should be quoted excluding GST in Figures and Words & free delivery should be made to the Surgical Store of this Hospital.

| Sl. No | Article Name | Accounting Unit | Rate/ Unit(Rs.) | GST % | MRP (Rs.) |
|--------|--|-----------------|-----------------|-------|-----------|
| 1. | Laminar Air Flow Hood Machine (Horizontal) Technical Specifications:- i) Air Flow Direction: Horizontal ii) Cabinet Fabricated in SS 304/316/MS Powder Coated Structure iii) Polycarbonate/Glass Slide Panel and Front Panel iv) Air velocity: 0.45 ± 0.05 mps (90 ± 20 FPM) Below six inch of Grill v) Miniplate HEPA filters H-14 99.999% Efficiency $0.3 \mu\text{m}$ vi) Pre-filter Glass Fiber Media EU-6 98% Efficiency $5 \mu\text{m}$ vii) Magnetic Differential pressure gauge viii) Centrifugal or Backward Curved type Air blower with Speed Variable ix) SS 304 working platform x) Germicidal Ultraviolet fluorescent light xi) Feather Touch controller for Blower/Light /Pressure Display xii) DOP Port/Atmospheric Nozzles xiii) Sound Level Minimum 65db on Scale xiv) Power supply Single Phase 220V 50HZ xv) Digital Pressure Gauge with Alarm xvi) Hour Meter | | | | |

[Handwritten signature]
27/2/23

Medical Superintendent-cum-Vice Principal
IPGME&R SSKM Hospital, Kolkata-700020

Memo. No. IPGME&R SSKMH/MSVP/1535/2023

Dated: 1/3/23

Copy to:

1. Mr. Atal Saha, Asst. Superintendent, SSKM Hospital, Kolkata for uploading in IPGME&R website
2. Notice Board, IPGME&R SSKM Hospital
3. Office Copy

[Handwritten signature]
27/2/23

Medical Superintendent-cum-Vice Principal
IPGME&R SSKM Hospital, Kolkata-700020