## GOVT. OF WEST BENGAL Office of the Medical Superintendent-cum-Vice Principal **IPGMER SSKM Hospital**

Memo	No	<b>IPGMER</b>	CEVARLI	JAACUM!	100%
MICHIO.	IVO.	IPGIVIER	22KIVIH	/IVI3VP/	1209

Dated: 17/02/25

## **QUOTATION NOTICE**

Sealed QUOTATION are invited from the reputed Companies/ Direct Importers/ Authorized Distributors/Traders for supply of the following items as mentioned below for use of Cytology Lab of IPGME&R-SSKM HOSPITAL, Kolkata-20

will submit the quotation in their own official pad along with the self attested copies of the following essential documents.

- 1. Valid Trade License/Enlistment
- 2. PAN Card of the Bidder/Bidder Company
- **GST Registration Certificate** 3.
- Last GST return of current year

The undersigned reserves the right to accept or cancel the whole procedure at any time without showing any reason.

- 1. Rate should be quoted (Figures and Words) & free delivery to the Pathology Store of this Hospital.
- 2. If the Schedule item is approved by CMS in that case purchase will be made from CMS approved firm
- Rate should be mentioned without GST, GST as applicable.
- 4. Tender committee reserves the right to decleare the Firm/ Company Blacklisted due to the falliure in supply within the stipulated period of 30 days each for two occasions during the quotation period.
- 5. Rate is Valid for one years or up to purchase below 1 lakhs.
- Batch test Report should be submitted at the time of Delivery of the item

SL.NO.	NAME OE THE ITEM	Packaging Unit	Rate Per Pk.
1.	COVID-19 RT PCR Kit	100 Rxn	> > > > + + + + + + + + + + + + + + + +
	Specification:-		
`	The kit is Compliance to Medical Device Rule (MDR) 2017 as amended till date.		
ě	<ol> <li>Availability of valid medical device license (CDSCO License) for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date</li> </ol>		
	<ol> <li>Kit should be ICMR approved</li> <li>The kit must be single tube assay for qualitative detection and differentiation</li> <li>The kit must be compatible with BIO-RAD CFX 96 Real Time PCR</li> </ol>		
	6. The kit should be consisting of atleast two target gene with endogenous IC		
	<ul><li>7. The kit should not cross react with other similar symptom causing viruses</li><li>8. Minimum detection limit is 10 copies/ml</li></ul>		
2.	Viral RNA Extraction kit	100Rxn	
	<ol> <li>The extraction kit should be able to extract viral RNA or total nucleic acid</li> <li>The extraction kit should be compatible with the following Human sample types; Respiratory specimens (nasopharyngeal or oropharyngeal swabs in VTM, bronchoalveolar lavage, tracheal aspirates, sputum, nasopharyngeal or oropharyngeal aspirates, and), Plasma, serum, urine, stool, cerebrospinal fluid (CSF), Urine sample, Rectal and Genital swab, Cell culture supernatants.</li> <li>The lysis and washing buffers should have room temperature storage for</li> </ol>		

**GST AS APPLICABLE** 

Medical Superintendent-cum-Vice Principal IPGME&R- SSKM Hospital, Kolkata-700020

Memo. No. IPGMER SSKMH/MSVP/ 1804

## Copy to:

- 1. IT Cell IPGME&R SSKM Hospital
- 2. Notice Board, IPGME&R SSKM Hospital
- Office Copy

Medical Superintendent-cum-Vice Principal GME&R- SSKM Hospital, Kolkata-700020