



Government of West Bengal
Office of The Superintendent
Sambhunath Pandit Hospital
Kolkata-700020

Memo No. SNPH/ 1429

Date :- 17-02-2026

Quotation Inviting Notice

Sealed quotations are hereby invited by the Superintendent, Sambhunath Pandit Hospital, Kolkata-20, from the reputed, Licensed eligible firms/ agencies/ individuals those who are Licensed to supply the following Equipments items/ Medicine at Sambhunath Pandit Hospital, Kolkata-20. Rate should be quoted including all charges in legible manner in figure & word as follows in separate financial bid therein:

Sl No	Name of the Medical Device with specification and standard	Financial Bid			Technical Bid
		Unit	Rate excluding of GST(INR)	GST Amount (INR)	
1.	BIPAP with humidifier (Specification details attach with quotation paper.)				It should contain- 1) Self-attested PAN Card & I.T. Return for the A.Y. 2025-2026. 2) Valid Drug License 3) Valid GSTIN & upto date return, 4) Valid Bank Details, 5) Valid License issued by the CDSCO/SDSCO in respect of the Medical Device/Equipment alongwith authorization from the respective CDSCO/SDSCO Licensed Holder.

The Sealed quotations must be dropped at the Office of the Superintendent, SNPH on and from 17-02-2026 within 24-02-2026 up to 02:00 P.M. Quotation will be opened as on 24-02-2026 at 03:00 P.M. at the office chamber of the undersigned.

Please follow the Two-Bid system mandatorily.

- Demonstration is mandatory within next 2 working days at the concerned Department, after opening Technical Bid, failing which the quotation will be treated as cancelled.
- It shall be the discretion of the Authority to get the quality & standards of the item verified from the respective Department prior to opening of financial bids and if so found not fit for use financial bid in respect of the tenderer shall be technically disqualified without any further communication.
- The rate should be quoted as per specification.
- Item name, Unit & Specification should not be changed.
- Tender/Quotation must contain two separate sealed envelopes for Technical Bid and Financial Bid respectively. The Financial Bid shall be opened only if the tenderer satisfies technical Qualification.
- Rate should not exceed by the rate notified by the NPPA & MRP.
- The bidder must enclose a copy of the License to Manufacture for Sale or Distribution of Medical Device/ The Registration Certificate to Sale or Distribute or offer for sale or deliver Medical Devices issued by the CDSCO/SDSCO in accordance with the Medical Device Rules, 2017.
- The Purchasing Committee has every right to accept/ reject any quotation without assigning any reason thereof.
- No other charges except the GST will be acceptable.
- Goods must be delivered within 7 (seven) days from the date of supply / work order failing which the order may be treated as cancelled and shall prevent to participate in Quotation/tender further for this Hospital.
- The quotation should be valid for **365 days** from the date of opening of quotation or till the procurement amount when reaches up to the maximum limit (Rs. 1 lac), whichever is earlier.
- If the L1 bidder found not able to comply the order within the stipulated time mentioned above, then the order will be automatically treated as cancelled. The undersigned will have the right to place the same before the L2 bidder without any communication to the L1 bidder.
- If any dispute OR reaction after using of the item arises, the selected bidder will have to replace the same without charging any cost and within next 03(three) days.
- **Lowest bid is not the sole criteria for selection; quality of article and technical standard mentioned above must be fulfilled.**
- Bill amount will be paid only after availability of allotment of fund in IFMS.


Superintendent

Sambhunath Pandit Hospital
Kolkata-700020

Date :- 17-02-2026

Memo No. SNPH/ 1429/1(B)

Copy forwarded to, for wide publicity and taking necessary action to :

1. The IT Cell Coordinator, Health & Family Welfare Department, Swasthya Bhaban, Kolkata with request to publish the notice in website of the Administrative Department (www.wbhealth.gov.in),
2. Notice Board
3. Office copy


Superintendent

Sambhunath Pandit Hospital
Kolkata-700020

(1)

Technical Specifications of BIPAP with humidifier:-

1. Should have all the modes of Bilevel ventilation S, T and ST modes,
2. Should be light weight and portable.
3. Should display device functions.
4. Should have leak compensation.
5. Unit should be USFDA or European CE certified from a 4 digit notified body and an applicable BIS Certification to be submitted if any.
6. Should have IPAP & EPAP ranging from Pressure 3-25 cm H2O.
7. Should have RAMP.
8. Should have low sound level.
9. Device should be able to compensate variable resistance related to different masks Nasal/oral face mask.
10. Should be able to detect and respond to Apnea and hypopnea.
11. Should have mask Fit/mask check alert.
12. Compliance data should be stored on internal or external memory devices which can be accessed through readers provided by the manufacturer.
13. Should be able to operate on 100V-240V AC.
14. Should be provided with integrated heated humidifier.
15. Manufacturer should have a wide national network with complete service, repair and application backup capability for the quoted equipment.
16. **Scope of Supply:** Main unit with 2 years warranty.
17. 1 reusable compatible reusable circuit, 1 reusable mask are also supply FREE OF COST with main machine.
18. Should provided software for downloading compliance date from the equipment at the hospital.
19. The model should be latest in series launched by the manufacturing company (certificate of the same should be provided by the manufacturer).
20. **Valid CDSCO certificate in name of Importer/Subsidiary/ OEM. Bidder CDSCO not valid until it is importer for quoted model.**

Yogiraj Ray

12/2/2026
Dr. Yogiraj Ray, MBBS, MD, DM
Associate Professor
Infectious Diseases
SNP Hospital (Annex-II), IPGME & R-SSKM, Kolkata
Reg. No-60194 (WBMC)