



Government of West Bengal  
Department of Microbiology

Institute of Post-Graduate Medical Education & Research

244, A.J.C. Bose Road, Kolkata - 700020


Ph : +9133-22236242 / 22231615 / 22239735 Extn. 2405



Memo no: /IPGME&R/Memo/346/23

Date: 19/6/2023

To  
The Director  
IPGME&R, Center of Excellence  
Kolkata-700020



Subject: Permission for inviting quotation for the procurement of Neurovirology Panel kit by Real Time PCR Based Detection under VRDL Project, Department of Microbiology, IPGME&R Kolkata  
Viral Research and Diagnostic Lab (VRDL) (Project No. V. 25011/171/2015-HR date 11th Feb 2016)

Dear Sir,

This is to inform you that there is an urgent requirement of a Neurovirology Panel kit by Real time PCR at the Department of Microbiology, IPGME&R, Kolkata for patient service. On average, 1kit of 24 rxns of Neurovirology Panel kit are required for testing the patient's sample. This is the **second** purchase of a Neurovirology Panel kit. For this, we require reputed companies/direct importers/authorized distributors to submit their respective quotations, following which the kit will be procured from chosen bidder.

**The approximately price of these kit (24 rxns) will be Rs. 90,000 approximately including GST.**

No	Item	Pack size	Qty	Approx Price
1	Neurovirology Panel kit by RT PCR	24 rxns	1	Rs. 90000 Including GST

This is for your kind attention and we request you to take necessary action.

Thanking You,

Regards

*Raja Roy*  
19/6/2023

Prof. (Dr.) Raja Roy  
HOD & PI (VRDL)

Prof. Raja Roy, HOD  
Department of Microbiology  
IPGME & R, SSKM Hospital  
Kolkata-20

Kolkata -700020

Memo no: 3701

Dated: 21/06/23

**QUOTATION NOTICE**

Sealed quotations (valid for a year from the date of submission) are invited from reputed companies/direct importers/authorized distributor firms for supply of the following items as mentioned below for use of Viral Research and Diagnostic Lab (VRDL) (Project No. V. 25011/171/2015-HR date 11th Feb 2016), under Department of Microbiology. IPGME&R, Kolkata. The last date of submission of quotation is 30/06/2023 at 1 P.M. The bidders will submit the quotation in their official pad along with the following essential documents:

1. Valid Trade License/Enlistment.
2. PAN card of the Bidder/Bidder Company.
3. GST registration certificate proof
4. Manufacturing License (in the case of the manufacturer).
5. GST Registration certificate.
6. Last Quarter's GST return.
7. IT Return of Last year.

The submitted quotation is to be kept at the office chamber of the undersigned on all working days from 03/07/2023 (Physical verification of samples and/or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patient since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof. The undersigned reserves the right to accept or cancel the whole procedure at any time without showing

Sl No	Name of the item	Specification.
1	Neurovirology Panel kit by Real Time PCR	24 tests

Director,  
IPGME&R, Kolkata.

  
19/06/23  
**Director**  
**I.P.G.M.E. & R.**  
Kolkata-700020

Memo No/Inst/3701/145 Dated 21/06/23

Copy forwarded for circulation to the:-

1. In-Charge, VRDL Project, Prof. Raja Ray
2. H.O.D, Dept. of Microbiology, Prof. Raja Ray
3. Accounts Officers
4. M.O.I.C (Project)
5. Atal Saha, Assistant Super for uploading the quotation to the website

  
21/06/23  
**Director**  
**I.P.G.M.E. & R.**  
Kolkata-700020

**Specification of Neurovirology panel detection kit by Real Time PCR & total Nucleic Acid Extraction kit**

<b>Product name</b>	<b>Specification.</b>
Neurovirology panel detection kit by Real Time PCR	<ol style="list-style-type: none"><li>1. The kit must be CE&amp;IVD approved</li><li>2. The kit must be having qualitative detection and differentiation</li><li>3. The kit must be compatible with BIO-RAD CFX 96 Real Time PCR</li><li>4. The kit should be consists of atleast HSV 1 &amp; 2, Epstein-Barr Virus and Varicella - zoster virus</li><li>5. Endogenous IC should be included in the kit</li><li>6. The kit should not cross react with other similar symptom causing viruses</li><li>7. The kit should be in 24 rxn format</li></ol>

*Raja Ray*  
19/6/2023

**Prof. Raja Ray, HOD**  
Department of Microbiology  
IPGME & R, SSKM Hospital  
Kolkata-20