

~~GOVT. OF WEST BENGAL~~
Office of the Medical Superintendent-cum-Vice Principal
IPGMER SSKM Hospital
Centre of Excellence
244 AJC Bose Road, Kolkata-700020

Memo. No. IPGMER SSKMH/MSVP/7763

Dated: 22.07.24

QUOTATION NOTICE

Sealed quotation are invited from the reputed Companies/ Direct Importers/ Authorized Distributors/Traders for supply of the following items as mentioned below for use of Microbiology Lab

The last date of submission of quotation on 29/7/24 and the same will be opened on the same date at 2.00 P.M. The bidders will submit the quotation in their own official pad along with the self attested copies of the following essential documents.

1. Valid Trade License/Enlistment
2. -PAN Card of the Bidder/ Bidder Company
3. GST Registration Certificate
4. Last GST return of current year

The quotation is to be submitted in the Pathology Store on all working days from 22/7/24 to 29/7/24 from 11 A.M to 4 P.M except 29/7/24 with in 2 P.M. . After taking receiving number from received section of MSVP Office The undersigned reserves the right to accept or cancel the whole procedure at any time without showing any reason. Rate should be quoted (Figures and Words) & free delivery to the Pathology Store of this Hospital.

SL.NO.	NAME OF THE ITEM	Pack Size	Maximum retail price	Rate per pack
1	Acanthameoba Detection Kit By Real Time PCR Specification:- 1. The kit must be having qualitative or quantitative acanthameoba detection 2. The Kit must be compatible with BIO-RAD CFX 96 Real Time PCR 3. Endogenous IC Should be included in the kit 4. The kit should not cross react with other similar symptom causing microorganisms. 5. The Kit should be in 25rxn format	25rxn		
2	Acanthameoba DNA Extraction Kit. Specification:- 1. The Purification kit should be able to extract Acanthameoba DNA 2. The purification kit should be compatible with the following sample types Human cerebrospinal fluid (CSF), Tissue 3. Spin Column based rapid and cost effective small scale extraction of high quality nucleic acid 4. The lysis and washing buffers should have room temperature storage for easy storing 5. The kit should be in 25rxn format	25rxn		

Handwritten signature and date: 22/7/24

****Sample must be approved by the demanding unit other wise no GRN will be done****

Medical Superintendent-cum-Vice Principal
IPGME&R- SSKM Hospital, Kolkata-700020

Memo. No. IPGMER SSKMH/MSVP/7763

Dated: 22.07.24

Copy to:

1. IT Cell IPGME&R SSKM Hospital
2. Notice Board, IPGME&R SSKM Hospital
3. Office Copy

Handwritten signature and date: 22/7/24

Medical Superintendent-cum-Vice Principal
IPGME&R- SSKM Hospital, Kolkata-700020