



Government of West Bengal

Office of the Medical Superintendent-cum-Vice Principal

IPGME&R SSKM Hospital, 244 AJC Bose Road, Kolkata-700020

EPABX: 22041100; E-Mail: msvpsskm@gmail.com; Office: 22231589; FAX: 22234658

Memo. No. IPGME&R SSKMH/MSVP/9100/2024

Dated: 28.08.24

QUOTATION NOTICE

Sealed quotations are invited from the reputed Companies/ Direct Importers/ Authorized Distributors/ Traders for supply of the following items as mentioned below for use at Urology Department of IPGME&R SSKM Hospital.

The last date of submission of quotation is on 04/09/24 and the same will be opened on the same date at 2.30 P.M. in presence of interested bidders, if any. The bidders will submit the quotation in their own official pad along with the self attested copies of the following essential documents.

1. Valid Trade License/Enlistment
2. PAN Card of the Bidder/ Bidder Company
3. GST Registration Certificate

The quotation is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from from 10.00 AM to 2.00 PM.

The undersigned reserves the right to accept or cancel the whole procedure at any time without showing any reason.

Rate should be quoted excluding GST in Figures and Words & free delivery should be made to the Surgical Store of this Hospital.

Sl. No	Article Name	Accounting Unit	Rate/Unit(Rs.)	GST %	MRP (Rs.)
1.	Eye Piece with Attachment Connector Specification: <ul style="list-style-type: none">• Should be universal to any Imaging System• Should have in Built Moire Filter to reduce honeycomb effect• Eye piece should be angled ocular• Ergonomic Handle ensures secure and comfortable handling during ureteroscopes	Each			
2.	Insert with working Channel Specification: <ul style="list-style-type: none">• Automatic tip design• Anti-reflection Tip design: The Anti-reflection Tip design prevents internal reflections	Each			

	<p>when putting an object right in front of the optical system</p> <ul style="list-style-type: none"> • Long lasting autoclavability • Size of working channel should be 4-5Fr • Working length should be 430mm • 7° direction of view 				
3.	<p>Image Fiber & light Fiber Specification:</p> <ul style="list-style-type: none"> • Image fibre should carry the image and light fibre should transmit the light evenly. • The large fibre bundle enables high resolution and an increased image size of the dual channel ureteroscope 	Each			
4.	<p>Luer Lock Attachment with Accessories Specification:</p> <ul style="list-style-type: none"> • Detachable Luer Lock Attachment to prevent back flow of water • New self-sealing single-use valves 	Each			

Technical Specifications of the Items are as submitted by the HOD, Department of Urology.

DD
28/8/24

Medical Superintendent-cum-Vice Principal
IPGME&R SSKM Hospital, Kolkata-700020

Memo. No. IPGME&R SSKMH/MSVP/9100/2024

Dated: 28.08.24

Copy to:

1. Mr. Atal Saha, Assistant Superintendent (NM), IPGME&R SSKM Hospital, Kolkata for uploading in IPGME&R website & West Bengal Health Portal
2. Notice Board, IPGME&R SSKM Hospital
3. Office Copy

DD
28/8/24

Medical Superintendent-cum-Vice Principal
IPGME&R SSKM Hospital, Kolkata-700020