

Government of West Bengal Office of the Director Institute of Post Graduate Medical Education and Research & SSKM Hospital Centre of Excellence

244, AJC Bose Road, Kolkata-700020

Memo No: Inst./

Dated.....

QUOTATION NOTICE

Sealed QUOTATIONS are invited from the reputed Companies/ Direct Importers/ Authorized Distributors/Traders for supply of the following items as mentioned below for Department of Physiology, IPGME&R, Kolkata-20.

Quotation will be done in two bid system:

- 1. Technical BID It shall include documents as stated under the Essential Documents heading.
- 2. Financial BID It shall include the rate of the mentioned item below in the figure and in words excluding GST.

The last date of submission of quotation is on 06/02/2025 and the same will be opened on 08/02/2025 at 11.00 A.M. in the presence of interested bidders, if any. The bidders will submit the quotation on their own official pad along with the following

Essential Documents:

- Valid Trade License
- 2. PAN Card of the Bidder/ Bidder Company
- 3. GST Registration Certificate
- 4. Compliance of Specification (in Office Letter head)

All the technically qualified bidders (to be notified later) should undergo demonstration in front of HOD, Department of Physiology, IPGME&R, Kolkata-20 before the opening of Financial BID.

The quotation is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 30 / 01 /2025 to 06 / 02 /2025 up to 2p.m.

The undersigned reserves the right to accept or cancel the whole procedure at any time without showing any reason.

- 1. Rate should be quoted (Figures and Words) & free delivery to the IPGME&R Store of IPGME&R and SSKM Hospital. Offer rate valid for 365 days.
- 2. If the schedule item is approved by C.M.S in that case purchase will be made from C.M.S approved firm.

SL No.	Name of Item	Specification	Rate per pc (Excluding GST) (Rs.)
1	Spirometer with Flowmeter of Pneumotach	Annexure – I (Attached herewith)	

Sd/-Director IPGME&R, Kolkata - 20

Memo No. Inst./ 848/1(4)

Dated: 30/01/2025

Copy forwarded for information & necessary action to:

MSVP, IPGME&R – SSKM Hospital, Kolkata.

- 2. Atal Saha, Asstt. Superintendent, IPGMER-SSKM Hospital for uploading in the IPGMER website and WB HEALTH website.
- 3. NOTICE BOARD, IPGME&R, Kolkata-700020.
- 4. Office Copy.

IPGME&R, Kolkata - 20

ar in grant	cification of Spirometer with Flowmeter of Pneumotach (141/s or more with accuracy 3 AIR- Ultrasonic Flow Sensor)	
Nn.	Specification	
1	Conformity to Standards American Thoracic society (ATS)/ European Respiratory Society (ERS) standards	
9 = 0		
1	Conformity to Safety and Quality Standards : USFDA	
11.7	FEV1.FEF50,FIF 50,FEF75,MVV, FET & MTT), SVC, Maximum ventilation volume	
1	(MVV)inspiratory capacity, Expiratory Reserve volume.	
	Should have Pre & Post Broncho dilatation Comparison.	
	Broncho provocation test is needed.	
6	Flowmeter: Pneumotach (14 l/s or more with accuracy 3 %).	
7	Level of Resistance should be less than 1.5 cm H2O/L/sec.	
8	Parameters should be measured with highest accuracy & reproducibility and should no	
	affected with high surrounding temperature and humidity levels.	
9	Should have Electronic Barometer & temperature sensors, for Automatic BTPS	
	correction.	
10	Overlaying of previous test curves for comparison.	
11.	Should have Real Time Flow Volume and Volume –time Time Traces on Computer screen.	
12	Capability to select and modify / add predicted equations.	
13	3-litre precision calibration syringe, standard accessories and manual is compulsory.	
14	Facility to interface for desktop/laptop computer is mandatory.	
14	System Should be supplied with Computer Interfacing package, Cables, Software should be supplied.	
16.	3-Litre Precision Calibration Syringe, Standard accessories &Manual should be supplied.	
17.	Availability of Disposable Mouthpieces : 500 Nos.	
18	Availability of Pulmonary Filters (100 Nos) with machine	
19,	Spare Pneumotach screens supplied : 1 No.	
.10	Availability of Nose clips 20 nos and Thermal paper roll-20 nos	
21.	Should have a digital display of all graphs and parameters and an alphanumeric keypad	
2.2	Overall dimensions in mm x mm x mm: 87X155X36	
21	Weight: 0.356kg	
24	Instruction Manuals: 3 set.	
25.	Warranty : 3Years.	
26.	Should have test report from central govt/ NABL/ILAC accredited lab covering all parameters.	
27.	Printing options – Direct to printer or with Easy One connect.	
28	Data Management - Easy One Connect HL7, XML, GDT with software.	
29	Data links – USB, Bluetooth up to 10000 tests	
30.	Age range - Spirometry > 4 years.	
31.	Dimensions – 87x155x36mm (HxBxT) 356g.	
12	Device Classifications – Type BF applied part.	
11.	Operating Conditions - Temp 0-40 degree C Rel. Humidity 5-90%Athmosph. Pressure 700-1060hPa.	
14.	Power supply – 5VDC, Standby 0.3W.	
15.	Exchangeable, 3.6 VDC	

Yours sincgrely,

医神经球球球球球球球球球球球球球 计下下中枢线运输工厂设计电话 医医疗证明 法人父的现在分词 下,人名《文

-

Prof. (Dr.) Arnab Sengupta

Professor and Head

Professor & Head Department of Physiology Department Of Physiology IPGME&R, Kolkata