

## GOVERNMENT OF WEST BENGAL OFFICE OF THE DIRECTOR INSTITUTE OF POST GRADUATE MEDICAL EDUCATION & RESEARCH 244, A.J.C BOSE ROAD, KOLKATA- 700020

Memo No. Inst/5089

Dated: 28, 9,22

## NOTICE

Acceptance of the lowest rate is not obligatory for the undersigned that reserves the right to accept or reject any offer without assigning any reason whatsoever. Rate must be submitted with company name & cat no. after calculating GST or other Taxes, if any according to the following format.

No:	Item	Make	HSN:	Pack Size	Qty
1.	HOT PLATE STIRRER	Borosil	8419	2L(HLS200),1NO	1

## Technical Specification for Magnetic stirrer with Hot plate

- a) Maximum stirring speed- upto 1500 rpm with stepless speed control and good speed stability
- b) Heating Capacity 300W
- c) Top plate Size -150x150mm
- d) Maximum capacity -2L
- e) Temperature range- up to 300 °C
- f) ) Temperature accuracy- +\_1° C
- g)Temperature control-PID and digital energy regulatory mode
- h) Hotplate-Should be chemically resistant to acid and alkali
- i) Top Plate Material Stainless steel
- j) Controls for both hotplate and stirrer should be provided with suitable indicators.
- k) Necessary electrical cables and magnetic beads should be provided.
- I) Warranty 1 y

## Copy forwarded for information:

- 1. Director, IPGME&R, Kolkata
- 2. Nodal Officer. Prof. Simanti Datta
- 3. Chairperson MRU Core Committee
- 4. Accounts Officer
- 5. Notice Board
- 6. Mr. Atal Saha, Asst. Supdt., for uploading the said order in the website of IPGMER

Director
IPGMER, Kolkata

Director P.O.M.S.S.P. Giller G. 1920 Medical Officer In-Charge

Store (Projects)